

## EU Declaration of conformity no. 200828-025

Product Name:	Vacuum Splint Kit AS 190 FW
Intended use:	The Vacuum Splint Kit are primarily intended to be used in prehospital and hospital environment, by professional trained emergency and hospital personnel to safely stabilise injured patient extremities during transport. The Vacuum Splints are suited for fixation of patients with hand, arm, ankle, leg injuries.  Including: Vacuum Splint AS 100 Vacuum Splint AS 120 Vacuum Splint AS 140 Vacuum Hand pump Repair Kit Bag for Splints Directions for use
SRN:	SE-MF-00003932
Basic UDI-DI: UDI DI: Germa Article No:	735001959P02VACSPLIGQ 07350019591925 23005002210
Manufacturer: Visiting address: Phone: Email: Web:	AB Germa Industrigatan 54-56, SE-29136 Kristianstad +46 (0)44 123030 info@germa.se www.germa.se
UKCA Representative  UK CA	Ferno (UK) Limited Ferno House, Stubs Beck Lane, Cleckheaton, West Yorkshire BD19 4TZ, England Telephone: +44 (0) 1274 851999, www.ferno.co.uk
Product class:	Class I according to rule 1 in Annex VIII in MDR 2017/745
Conformity procedure:	Self-certification according to Annex IV in MDR 2017/745
Identification:	All products with serial numbers issued from; LOT number: 517470 Date: 2021-05-25 (yyyy-mm-dd).

## laration statement;

This EU declaration of conformity is issued under the sole responsibility of the manufacturer AB Germa. The devices covered by this declaration is in conformity with the requirements in the European Medical Device Regulation 2017/745.

Signed in Kristianstad, Sweden on behalf of AB Germa by:

Position: Managing Director Name: Björn Holmqvist

Date: 2021-05-25

Postadress/Postal address Adress/Goods address

Telefon/Phone

Telefax/Fax

E-mailWeb page

Bankgiro Bank account

SWEDEN